

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
1	Black Silk Braided Size 1, ½ Circle RB e Needle All Sizes Taper Point														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glylsilk	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	Thread break	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SILK	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Popular International, Multan	Covidien/ USA	SOFSILK	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Silk Braided (STH202)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Black Braided Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak needle	Non Responsive
2	Black Silk Braided Size 2, Curved Cutting e Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glylsilk	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
2	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Silk Braided (STH201)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Black Braided Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
3	Black Silk Braided Size 2/0, ½ Circle RB, Curved / Reverse / Straight Cutting e Needle All Sizes													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	GlysilK	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SILK (30mm)(60mm)	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
3	M/S Popular International, Multan	Covidien/ USA	SOF SILK	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Silk Braided (ST 207) (ST 204) (SR 104) (SS 004)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Black Braided Silk (30mm)(25mm,30	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	needle break Non Responsive
4	Black Silk Braided Size 3/0, ½ Circle RB, Curved Cutting e Needle All Sizes													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	GlysilK	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SILK	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
3	M/S Popular International, Multan	Covidien/ USA	SOF SILK	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Silk Braided (ST 210) (ST 209) (SI 301)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Black Braided Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Needle blend Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
5	Black Silk Braided Size 4/0, Curved Cutting e Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glysilik	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
2	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Silk Braided (SI 104) (SI 302)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Black Braided Silk	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
6	Polymide (Nylon) Non Absorbable + Spatulated Double Needle Size 10/0														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Daclon, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Ethilon	Y	Y	Y	Y	Y	Y	Y	Y	N	No sample	NO SAMPLE	Non Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Nylon Monofilament Sutures	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
7	Catgut Chromic Size 1, ½ Circle RB / Taper Point e Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Chromic Catgut	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread break	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme GUT	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (CT 205) (CT 207)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Chromic Catgut	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	needle weak	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
8	Catgut Chromic Size 2, ½ Circle RB / Taper Point ě Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Chromic Catgut	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread break	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme GUT	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (CTH 201)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Chromic Catgut	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	needle break	Non Responsive
9	Catgut Chromic Size 2/0, ½ Circle RB / Taper Point ě Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Chromic Catgut	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread break	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme GUT	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (CT220) (CT221) (CT218)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Chromic Catgut	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak Swag	Non Responsive
10	Catgut Chromic Size 4/0, ½ Curved Needle RB / Taper Point ě Needle All Sizes														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Chromic Catgut	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak swag	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme GUT	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (CT224)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Chromic Catgut	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread break	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
11	Polyglactin (Anti Bacterial) Rx Vicryl 0 (1/3 Circle Needle)													
1	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl Plus	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive(J needle required)
2	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic Acid 910	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak swag, not anti-bacterial	Non Responsive
12	Polyglactin (Anti Bacterial) Rx Vicryl 1 (1/3 Circle Needle)													
1	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl Plus	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic Acid 910	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	not anti-bacterial	Non Responsive
13	Polyglactin (Anti Bacterial) Rx Vicryl 2/0													
1	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl Plus	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic Acid 910	Y	Y	Y	Y	Y	Y	Y	N	No Sample	No Sample	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
14	Polyglactin / Polyglycolic Acid / Lactomer Size 1, ½ Circle RB e Needle														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak Swag	Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Surgicryl 910 SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak Needle	Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SORB (90cm)(75cm)	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic Acid 910	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
15	Polyglactin / Polyglycolic Acid / Lactomer Size 2/0, ½ Circle RB e Needle														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	Swag Weak	Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Surgicryl 910 SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SORB	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
16	Polyglactin / Polyglycolic Acid / Lactomer Size 3/0, ½ Circle RB ñ Needle														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread weak	Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Surgicryl 910 SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SORB	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak swag	Non Responsive
17	Polyglactin / Polyglycolic Acid / Lactomer Size 4/0, ½ Circle RB ñ Needle														
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme SORB	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable-Vicryl	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak swag	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
18	Polyglactin / Polyglycolic Acid / Lactomer Size 6/0 (Ophthalmology) ē Needle All Sizes													
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Surgicryl 910 SMI	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Thread break easily	Non Responsive
2	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Thread break easily	Non Responsive
3	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable- Vicryl	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic Acid 6/0	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Needle Moldable	Non Responsive
19	Polyglactin / Polyglycolic Acid / Lactomer Size 5/0 RB ē Needle 17μm													
1	M/S Popular International, Multan	Covidien/ USA	POLYSORB	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Absorbable- Vicryl	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO PAG Polyglycolic Acid/PGLA Poly Lactic co Glycolic	Y	Y	Y	Y	Y	Y	Y	N	NO SAMPLE	NO SAMPLE	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
20	Polypropylene Knitted Non Absorbable / Absorbable Mesh Undyed All Sizes 15x15													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Klas Medikal, Turkey	Svpro Polypropylene Mesh	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Medicamp International, Rawalpindi	Kollsut USA FDA	Kollsut Sietka Mesh Polypropylene Knitted Non	Y	APPLY MI	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Medilution Health Care, Karachi	Beta Mesh Turkey		Y	APPLY MI	Y	Y	Y	Y	Y	N	NO SAMPLE	NO SAMPLE	Non Responsive
4	M/S Popular International, Multan	Covidien/ USA	PARIETENE PP 15x15 CM	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Mesh	Y	Y	Y	Y	Y	Y	Y	N	NO SAMPLE	NO SAMPLE	Non Responsive
21	Polypropylene Knitted Non Absorbable Mesh Undyed All Sizes 6 x 11													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Klas Medikal, Turkey	Svpro Polypropylene Mesh	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Medicamp International, Rawalpindi	Kollsut USA FDA	Kollsut Sietka Mesh Polypropylene Knitted Non	Y	APPLY MI	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Medilution Health Care, Karachi	Beta Mesh Turkey		Y	APPLY MI	Y	Y	Y	Y	Y	N	NO SAMPLE	NO SAMPLE	Non Responsive
4	M/S Popular International, Multan	Covidien/ USA	PARIETENE PP 11x06 CM	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Mesh	Y	Y	Y	Y	Y	Y	Y	N	NO SAMPLE	NO SAMPLE	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
22	Polypropylene Size 1, ½ Circle RB / Taper Point ÷ Needle All Sizes													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread weak Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polyproplene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak needle Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
4	M/S Popular International, Multan	Covidien/ USA	SURGI PRO	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
5	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (NT202)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
6	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil Responsive
7	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak Swag Non Responsive
23	Polypropylene Size 1, ½ Circle RB / Taper Point ÷ Double Needle All Sizes													
1	N.Q													

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
24	Polypropylene Size 2/0, ½ Circle RB / Taper Point ð Needle All Sizes													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surgigut Chromic (NT205) (NS002) (NT206) (NR302)	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
7	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
25	Polypropylene Size 2/0, ½ Circle / Taper Point ð Double Needle All Sizes													
1	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
26	Polypropylene Size 3/0, RB / Cutting / Taper Point e Needle All Sizes													
1	M/S K. S. Agencies Pharmaceuticals Distributors, Karachi	Huaiyin Medical Instruments Co Ltd, China	Glytin	Y	Apply MD	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
2	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
3	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surglene (NX302) (NX303)	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
7	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene (25mm,30mm)(16	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak thread	Non Responsive
27	Polypropylene Size 3/0, 26mm, ½ Circle / Taper Point e Double Needle , 90cm length													
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surglene (NTD203)	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample			
28	Polypropylene Size 4/0, RB / Curved Cutting / Taper Point e Needle All Sizes													
1	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE (20mm)(19mm)	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surglene (NX307) (NX308)	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	Weak Swag	Responsive
4	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
29	Polypropylene Size 4/0, 16mm, ½ Circle / Taper Point e Double Needle , 90cm length													
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surglene (NTD208)	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
30	Polypropylene Size 4/0, 20mm, ½ Circle / Taper Point ÷ Double Needle , 90cm length														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Vikor Health care, Karachi	Vikor Healthcare Pvt Ltd	Surgiline Surglene (NTD206)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
6	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
31	Polypropylene Size 5/0, RB / Curved Cutting / Taper Point ÷ Needle All Sizes														
1	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE (17mm 1/2)(15mm 3/8)	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Popular International, Multan	Covidien/ USA	SURGIPRO (16mm 17cm)(17mm 90cm)	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
32	Polypropylene Size 6/0, 13mm ½ C, Taper point, Double Needle														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
2	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
33	Polypropylene Size 7/0, RB / Cutting (W8702 / W8304) ē Needle All Sizes														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
2	M/S Popular International, Multan	Covidien/ USA	SURGIPRO	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	not as per specs	Non Responsive
3	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread weak	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
34	Polypropylene Size 7/0, 8mm/9.3mm ½ C, Taper point, Double Needle														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Polypropylene, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak needle	Non Responsive
2	M/S Sind Medical Store, Karachi	Deme TECH Corporation U.S.A	Deme LENE	Y	Apply MD	Y	Y	Y	Y	Y	Y	Y	REJECTED	needle bend	Non Responsive
3	M/S Popular International, Multan	Covidien/ USA	SURGI PRO	Y	MD	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Prolene	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
5	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Prolene Polypropylene	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	thread weak	Non Responsive
35	Surgical Steel Wires No. 4														
1	M/S Popular International, Multan	Covidien/ USA	STEEL	Y	N	Y	Y	Y	Y	Y	Y	N	NO Sample	NO Sample	Non Responsive
2	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Other	Y	Y	Y	Y	Y	Y	Y	Y	N	NO Sample	NO Sample	Non Responsive
3	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Stainless Steel Wire	Y	Y	Y	Y	Y	Y	Y	Y	N	NO Sample	NO Sample	Non Responsive

TECHNICAL COMPARITIVE STATEMENT FOR TENDER OF SUTURES FOR THE YEAR 2022-23

S.N	Name of Firm	Manufacturer	Brand Name	Compulsory Criteria								Sample (Approved/ Rejected)	REASON OF REJECTION	REMARKS	
				Agency Agreement (Y/N)	Registration / Provisional Registrati	Establishment License/ import license	Manufacturer specification	Valid Quality Certificate (Y/N)	Valid Free Sale Certificate (Y/N)	Affidavits (Y/N)	Sample				
36	Surgical Steel Wires No. 5														
1	M/S Anwar & Sons, Rawalpindi	SIM, Belgium	Steel, SMI	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	weak wire	Non Responsive
2	M/S Popular International, Multan	Covidien/ USA	STEEL	Y	N	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
3	M/S Muller & Phipps, Lahore	Johnson & Johnson	Sutures Non-Absorb- Other	Y	Y	Y	Y	Y	Y	Y	Y	Y	APPROVED	Nil	Responsive
4	M/S Akram & Brothers, Lahore	Foosin Medical Supplies Inc Ltd, China	WEGO Stainless Steel Wire	Y	Y	Y	Y	Y	Y	Y	Y	Y	REJECTED	No experience, not used	Non Responsive
37	Sterna Polyester														
1	N.Q														

Dr. Umair Ahmad
Associate Professor of Surgery

Mr. Khaqan Saleem
Pharmacist Cardiac Center

Miss. Beenish Rana
Pharmacist Kidney Center

Miss Amna Saeed
Pharmacist B.V. Hospital,

Dr. Shabana Anwar
Pharmacist Sadiq Abbasi Hospital,

Dr. Imran Nazir
Pharmacist A & E Department,
Member Sub Committee

Dr. Ahsan Firdous
A.M.S Store
Cardiac Center Bahawalpur

Dr. Syed Tariq Hassan
A.M.S Medical Store
Sadiq Abbasi Hospital
Bahawalpur

Dr. Sajjad Mozzam
A.M.S Medical Store
B.V. Hospital Bahawalpur

Dr. Wajhat Hussain
Demonstrator Community Medicine
Department, Bahawalpur

Dr. Zafar Iqbal Ghafoor Assistant
Professor of Pharmacology Deptt.

Dr. Sara Raza
Assistant Professor of Pathology Deptt.

Dr. Saliha Zafar
Consultant Pathologist Pathology Department

Prof. Dr. Ameer Ahmed Malik
Head of Pediatric Medicine Department

Prof. Dr. Muhammad Nasir Ali
Professor of Orthopedic Department
B.V. Hospital, Bahawalpur
(Chairman TAC)